

	Application No.	Applicant(s)
	10/761,943	PENG, HONG
Notice of Allowability	Examiner	Art Unit
	Michael P. Stafira	2877
The MAILING DATE of this communication apperation apperation allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI of the Office or upon petition by the applicant. See 37 CFR 1.313	ears on the cover sheet with the of (OR REMAINS) CLOSED in this ap or other appropriate communicatio GHTS. This application is subject	correspondence address oplication. If not included n will be mailed in due course. THIS
1. 🔀 This communication is responsive to amendment filed 4/24	<u>1/2006</u> .	
2. The allowed claim(s) is/are 24-45.		
Acknowledgment is made of a claim for foreign priority unally All b) □ Some* c) □ None of the: Certified copies of the priority documents have Certified copies of the priority documents have	been received. been received in Application No	
 Copies of the certified copies of the priority documents International Bureau (PCT Rule 17.2(a)). 	cuments have been received in this	national stage application from the
* Certified copies not received:		
Applicant has THREE MONTHS FROM THE "MAILING DATE" noted below. Failure to timely comply will result in ABANDONM THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.		complying with the requirements
4. A SUBSTITUTE OATH OR DECLARATION must be subm INFORMAL PATENT APPLICATION (PTO-152) which give		
5. CORRECTED DRAWINGS (as "replacement sheets") mus	st be submitted.	
(a) ☐ including changes required by the Notice of Draftspers	on's Patent Drawing Review (PTO	-948) attached
1) 🗌 hereto or 2) 🔲 to Paper No./Mail Date		
(b) ☐ including changes required by the attached Examiner's Paper No./Mail Date		
Identifying indicia such as the application number (see 37 CFR 1 each sheet. Replacement sheet(s) should be labeled as such in t		
6. DEPOSIT OF and/or INFORMATION about the deposit attached Examiner's comment regarding REQUIREMENT		
Attachment(s) 1. ☐ Notice of References Cited (PTO-892)	5 Notice of Informal I	Patent Application (PTO-152)
2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)	6. ☐ Interview Summary	• • • • • • • • • • • • • • • • • • • •
3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/0	Paper No./Mail Da	ate
Paper No./Mail Date4. ☐ Examiner's Comment Regarding Requirement for Deposit	8 ⊠ Evaminar's Statem	ent of Reasons for Allowance
of Biological Material	9. ☐ Other	ent of reasons for Allowance

DETAILED ACTION

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR
 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

The application has been amended as follows:

- a. In claim 35, delete "the" in front of "shaking platform" and insert --a--;
- b. In claim 44, delete "the" in front of "shaking platform" and insert --a--.

Allowable Subject Matter

- 2. Claims 24-45 are allowed over the prior art of record.
- 3. The following is an examiner's statement of reasons for allowance:

Regarding claim 24, the prior art fails to disclose or make obvious an apparatus for online monitoring the cell concentration of biological culture medium in a shaking environment of
an incubator/shaker having a probe fixture means for mounting and orientating at least one said
light emission source and at least one said photodetector in fixed positions relative to each other,
and for holding said container firmly without any relative movement with respect to said light
emission source, said photodetector and a shaking platform of said incubator/shaker during the
on-line monitoring process while said container and the probe fixture are under a continuously

Art Unit: 2877

shaking condition, and in combination with the other recited limitations of claim 24. Claims 25-34 are allowed by the virtue of dependency on the allowed claim 24.

Regarding claim 35, the prior art fails to disclose or make obvious a method for on-line monitoring the cell concentration of biological culture medium in a shaking environment of a incubator/shaker having the steps of preventing the relative movement of said container with respect to the position of the light emission source, the photodetector and the shaking platform of the incubator/shaker during the on-line monitoring process while said container and the probe fixture are under a continuously shaking condition, and in combination with the other recited limitations of claim 35. Claims 36-43 are allowed by the virtue of dependency on the allowed claim 35.

Regarding claim 44, the prior art fails to disclose or make obvious an apparatus for online monitoring the cell growth curve of biological cell culture in a shaking environment of an
incubator/shaker having a probe fixture means for mounting and orientating at least one said
light emission source and at least one said photodetector in fixed positions relative to each other,
and for holding said container firmly without any relative movement with respect to said light
emission source, said photodetector and the shaking platform of said incubator/shaker during the
on-line monitoring process while said container and the probe fixture are under a continuously
shaking condition, and in combination with the other recited limitations of claim 44. Claim 45 is
allowed by the virtue of dependency on the allowed claim 44.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue

Art Unit: 2877

fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael P. Stafira whose telephone number is 571-272-2430. The examiner can normally be reached on 4/10 Schedule Mon.-Thurs..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Gregory Toatley can be reached on 571-272-2800 ext. 77. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Michael P. Stafira Primary Examiner Art Unit 2877 Page 4

June 6, 2006